



USP <88> Class Plastic Designation

Sponsor: Garlock Sealing Technologies
1666 Division Street
Palmyra, NY 14522-9350

Attn: Matt Tones

Test Nos.: T04-2417, 2418, 2419, Amendment I

Test Material: GYLON® Style 3504 (blue)

This test material met the requirements of USP 27 - NF 22, General Chapter <88>, for Class VI designation at a temperature extraction of 121°C for 1 hour.

Jonia Schaubroeck

Study Director

3-10-09

Date

as per original release date

T54-8-10



Amendment of USP Class VI Designation Certificate
(Reference SRP-005)

Regarding Test Nos.: T04-2417, 2418, 2419 Amendment # 1

The following correction has been made to the above mentioned document:

Amendment 1: The USP general chapter "<88>" has been added to the title and statement section.

Reason: The USP general chapter governing USP Class Plastics Designation testing is chapter <88>. The chapter was added to the document for clarification purposes.

Jonia Schaubroeck

Study Director

3-10-09

Date

*as per original release date
754-840*