



USP <88> Class Plastic Designation

Sponsor: Garlock Sealing Technologies
1666 Division Street
Palmyra, NY 14522-9350

Attn: Matt Tones

Test No(s): T10-0691, 0692, 0693, Amendment I

Test Material: One (1) GYLON 3510

This test material meets the requirements for USP 27 – NF 22, General Chapter <88>, for Class VI designation at a temperature extraction of 121°C for 1 hour.

Sonia Schaubrock

Study Director

4-8-10

Date



Amendment of USP Class VI Designation Certificate
(Reference SRP-005)

Regarding Test Nos.: T10-0691, 0692, 0693

Amendment # 1

The following correction has been made to the above mentioned document:

Amendment 1: The USP – National Formulary editions (USP 27 – NF 22) and general chapter (<88>) have been added to the USP Class Plastics Designation Certificate originally issued by Tonia Schaubroeck on July 6, 2004.

Reason: The USP edition and general chapter governing USP Class Plastics Designation testing were not referenced on the original certificate. This information was added to the document for clarification purposes.

Tonia Schaubroeck

Study Director

4-8-10

Date